

EasyOne-CS

Field Service Manual



EasyOne-CS Spirometer

Manufactured for GE Medical Systems Information Technologies

ndd Medizintechnik AG
Technoparkstrasse 1
CH-8005 Zürich, Schweiz
www.ndd.ch

Contents

| | | |
|-----------|--|-----------|
| 1 | Manual Purpose | 3 |
| 2 | Intended Audience | 3 |
| 3 | Introduction | 3 |
| 4 | Operation Principle | 3 |
| 5 | Technical Details | 4 |
| 5.1 | EasyOne-CS Spirometer Handheld Unit..... | 4 |
| 5.2 | Screen-Connector..... | 5 |
| 5.3 | Block Diagram | 6 |
| 5.4 | Printed Circuit board..... | 6 |
| 6 | Specifications..... | 7 |
| 6.1 | EasyOne-CS Model 2001 Spirometer | 7 |
| 6.2 | EasyOne-CS Model 2010 Screen-Connector | 8 |
| 7 | Replacement Parts List..... | 8 |
| 7.1 | Manuals | 8 |
| 7.2 | Accessories..... | 9 |
| 7.3 | Spare Parts | 11 |
| 7.4 | Exchange Unit | 12 |
| 8 | Checkout Procedure | 13 |
| 9 | Hidden Keys | 14 |
| 10 | Device Error Codes | 15 |
| 11 | EasyOne-CS Device Software Update..... | 16 |
| 11.1 | Introduction | 16 |
| 11.2 | Validated Upgrade Steps | 16 |
| 11.3 | Preparation | 16 |
| 11.4 | Unpacking of Upgrade Package | 16 |
| 11.5 | Installation of EasyFlash PC Software..... | 16 |
| 11.6 | EasyOne-CS Device Software Update Procedure..... | 18 |
| 12 | EasyWare PC Software Installation..... | 19 |
| 13 | EasyOne-CS Lithium Battery Replacement | 21 |
| 13.1 | Background..... | 21 |
| 13.2 | Environment..... | 22 |
| 13.3 | Preparation (save database)..... | 22 |
| 13.4 | Battery exchange procedure | 23 |
| 13.5 | Restart EasyOne-CS | 24 |
| 13.6 | Restore database | 24 |
| 14 | EasyOne-CS board with sensor and housing parts Replacement..... | 24 |
| 14.1 | Background..... | 24 |
| 15 | Appendix..... | 26 |
| 15.1 | Schematic Diagrams..... | 26 |
| 15.2 | WEEE Selective Treatment | 33 |

Document History

| Version | Description | Datum |
|--------------------|--|------------------|
| 2030434-006 Rev. A | First Release | August 8, 2006 |
| 2030434-006 Rev. B | Chapter 7.1 with new content Chapter 7.2 with V2.9.5 Chapter 9 with additional comment for 1-2-5 Chapter 11.2 with Windows 2000, XP or Vista Chapter 11.3 update: added resource update Chapter 8 updated and renamed as 'Checkout Procedure' | August 31, 2009 |
| 2030434-006 Rev. C | Chapter 12 EasyWare PC Software Installation - new This page with additional addresses | Octobre 27, 2009 |
| 2030434-006 Rev. D | Header info modified to "EasyOne-CS Field Service Manual" Chapter 7.2 Accessories – CD pictures adapted to V2.9.5 Chapter 11 modified to document update FW v4.1 to v4.3 | November 6, 2009 |
| 2030434-006 Rev. E | Telephone number (US) changed Chapter 7.1: Chinese manual is no longer available New chapter 15.2 WEEE Selective Treatment | April 1, 2014 |

**Europe:**

ndd Medizintechnik AG
 Technoparkstrasse 1, CH-8005 Zurich, Switzerland
 Tel.: +41 44 445 29 70
 Fax: +41 44 445 25 31
 Web: www.ndd.ch

USA:

ndd Medical Technologies
 Two Dundee Park
 Andover, MA 01810
 Tel: +1 877 904 0090
 Fax: +1 978 470 0924
 Web: www.nddmed.com

Asia - Please contact GE Healthcare:

Asia Headquarters
 GE China Co., Ltd.
 No1 Huatuo Road,
 Zhangjiang Hi-Tech Park Pudong,
 Shanghai, P.R.China 201203
 Tel: +86 21 38777888
 Fax: +86 21 38777402

Distributed by:

GE Medical Systems Information Technologies, Inc.
 8200 West Tower Avenue, Milwaukee, WI 53223 USA
 Tel.: +1 414 355 5000
 +1 800 437 1171 (USA only)
 +1 800 668 0732 (Canada only)
 Fax: +1 414 355 3790

1 Manual Purpose

This manual supplies technical information for service representative and technical personnel so they can maintain the equipment to the assembly level. Use it as a guide for maintenance and electrical repairs considered field repairable. Where necessary the manual identifies additional sources of relevant information and or technical assistance.

See the operator manual for the instructions necessary to operate the equipment safely in accordance with its function and intended use.

2 Intended Audience

This manual is intended for the person who uses, maintains, or troubleshoots this equipment.

3 Introduction

The EasyOne-CS Spirometer uses digital ultrasonic flow measurement technology for fast, accurate, reliable operation. The EasyOne-CS was designed for full range testing in primary care, specialty physician, industrial and hospital settings. It offers automated comparison to predicted values, pre- and post Broncho Dilator comparisons, interpretation of results and test quality control that automatically assesses patient effort offering helpful suggestions to obtain optimum results. Optional easy to read and interpret color printout with graphic display for instant test review is available. The single-patient-use spirotte minimizes risk of cross-infection.

4 Operation Principle

This EasyOne-CS flow sensor is equipped with two ultrasonic transducers, mounted at different sides of the flow channel. In order to determine gas flow (and molar mass), ultrasonic pulses are transmitted in up- and downstream direction. Precisely measured transit-times are the basis for determining flow and molar mass. The following figure shows a diagram of the flow sensor including its internal sound transmission path.




Figure 1: Schematic of EasyOne-CS flow sensor

5 Technical Details

5.1 EasyOne-CS Spirometer Handheld Unit

The following pictures show the top, rear and bottom views of the EasyOne-CS.




Figure 2: Top view of EasyOne-CS



Figure 3: Rear view with empty battery compartment

The two screws on the back side are used to open EasyOne-CS. With the exception of the lithium battery there are no serviceable parts inside the unit. Therefore do not open the unit!



Figure 4: Bottom view of EasyOne-CS

The bottom view shows the name plate, the serial number and the EasyOne-CS-to-Screen-Connector.

5.2 Screen-Connector

The following pictures show the top and bottom views of the EasyOne-CS Screen-Connector.



Figure 5: Screen-Connector top view




Figure 6: Screen-Connector bottom view

The Screen-Connector uses a standard USB connection to the PC.

5.3 Block Diagram

The following picture shows the functional block diagram of the EasyOne-CS Spirometer.



EZI-BlockDiagram.sdr

5.4 Printed Circuit board

All electrical components of EasyOne-CS are mounted on a single PCB.

With the exception of the lithium battery no components can be serviced.

The following two pictures show a top and bottom view of the PCB.




Figure 7: Bottom view of EasyOne-CS PCB

The lithium battery located on the top right of the PCB has an approximate life time of 5 to 10 years. In early revisions of the hardware the lithium battery is soldered on the board.



Figure 8: Top view of EasyOne-CS PCB (without EMI shielding)

6 Specifications

6.1 EasyOne-CS Model 2001 Spirometer

| | |
|------------------------|--|
| Size: | 83 x 158 x 43 mm (3.3 x 6.2 x 1.7 inches) |
| Weight: | 242 grams (8.6 ounces), with batteries |
| Measuring accuracy | Volume: ±2% or 0.050 l Flow: ±2% or 0.020 l/s, (except PEF) PEF: ±5% or 0.200 l/s MVV: ±5% or 5 l/min. |
| Measuring range: | Volume: ±12 l Flow: ±16 l/s |
| Resistance: | below 0.3 cm H ₂ O/L/s |
| Display: | 64 x 160 graphic display |
| Data entry: | 14-key keyboard |
| Data memory: | For up to 700 tests, stored in EasyOne CS |
| Test modes: | FVC, FVL, Slow VC, MVV, Pre/Post |
| Parameters: | FVC, MVV, FEV6, FEV1, FEV1/FVC, FEV1/VC, FEV1/FEV6, MEF25 (FEF75), MEF50 (FEF50), MEF75 (FEF25), MEF25%-75%, PEF, FET, FIVC, PIF, IVC, ERV, IRV, pre-post % variation, Lung Age |
| Respiratory tube: | Disposable spirotte respiratory tube |
| Measurement principle: | Ultrasound transit-time measurement |
| Adult predicted: | NHANES-III, Knudson_83, Knudson_76, Crapo, Morris, ERS (ECCS/EGKS), Forche (Austria), Sapaldia (Switzerland), Roca (Spain), Berglund, Gulsvik, Hedenström, Gore, Cherniak (only for MVV) |
| Pediatric predicted: | Dockery (Harvard), Hsu, Zapletal, Polgar, Hibbert |
| Power supply | 2 alkaline batteries, Type AA, 1.5V |
| Power consumption: | Typical 0.6 W |
| Report: | A4 |
| Storage: | Temperature: -20 to 50 °C Relative humidity: 5% to 95% Ambient pressure: 500 to 1060 hPa |

| | |
|-------------------------------|--|
| Operating conditions: | Temperature: 0 to 40 °C Relative humidity: 0% to 95% Ambient pressure: 500 to 1060 hPa |
| Certifications and standards: | CE Declaration of Conformity C CSA US approval, CAN/CSA-C22.2 No. 601.1-M90, S1-94, CSA 601.1 Amendment 2:1998, UL Std No. 2601.1 FDA 510 (k) approval, K993921 EasyOne-CS meets or exceeds the published targets of the European Respiratory Society (ERS), the American Thoracic Society (ATS) and the National Lung Health Education Program (NLHEP). |
| Instrument classification: |  Type BF applied part (when used with a PC, protection class 2 according to IEC 60601-1) Instrument not suitable for use in flammable anesthetic gases in mixtures with O ₂ or NO. |
| Life time: | 7 years |

6.2 EasyOne-CS Model 2010 Screen-Connector

| | |
|-----------------------|--|
| Size: | 64 x 44 x 25 mm (2.5 x 1.7 x 1.0 inches) |
| Weight: | 82 grams (3 ounces) |
| Power supply: | From USB port |
| Power consumption: | Type 0.15W |
| Function: | Connects the EasyOne-CS spirometer to a PC |
| Interface: | Standard USB type A connector for connection to PC |
| |  |
| | Pins: 1 = VBus, 2 = D-, 3 = D+, 4 = GND |
| Storage: | Temperature: -20 to 50 °C Relative humidity: 5% to 95% Ambient pressure: 500 to 1060 hPa |
| Operating conditions: | Temperature: 0 to 40 °C Relative humidity: 0% to 95% Ambient pressure: 500 to 1060 hPa |
| PC: | The PC must comply with corresponding IEC standard (ex. IEC 60950-1. The user is responsible that requirements of IEC 60601-1-1 for safety of medical electrical systems are met.) |


7 Replacement Parts List

The following parts are available for replacement:

7.1 Manuals

| Part Number Short Description | Description |
|----------------------------------|--------------------------------------|
| 2030434-006 MNL SVCE EASYONE-CS | Field Service Manual (this document) |
| 2030434-013 | Operator's Manual German |
| 2030434-014 | Operator's Manual English |
| 2030434-015 | Operator's Manual French |
| 2030434-016 | Operator's Manual Italian |
| 2030434-017 | Operator's Manual Spanish |
| 2030434-018 | Operator's Manual Portuguese |
| 2030434-019 | Operator's Manual Danish |
| 2030434-020 | Operator's Manual Dutch |
| 2030434-023 | Operator's Manual English-US |



7.2 Accessories

| Part Number Short Description | Description |
|--------------------------------------|---|
| 2030431-003 CABLE EASYONE-CS - PC | Screen-Connector, connects EasyOne-CS with the PC via USB connection  |
| 2030431-005 CD EASYWARE V2.9.5 | CD with EasyWare V2.9.5 software  |
| 2030431-006 SPLY SPIRETTES #50 | 50 spirettes  |
| 2030431-007 SPLY SPIRETTES #200 | 200 spirettes  |


| Part Number Short Description | Description |
|---|------------------------|
| 2030431-008 SPLY NOSECLIP #25 | 25 nose clips |
|  | 100 pads for noseclips |
| 2030431-010 SPLY CALIBRATION SYRINGE EASYONE-CS | Calibration syringe |

7.3 Spare Parts

| Part Number Short Description | Description |
|--|--|
| 2033231-001 SPARE EASYONE-CS FRONTHALF | Case front half  |
| 2033231-002 SPARE EASYONE-CS REARHALF | Case rear half  |
| 2033231-003 SPARE EASYONE-CS BATTERY DOOR | Battery door  |
| 2033231-004 SPARE EASYONE-CS KEYPAD | Keypad  |

| Part Number Short Description | Description |
|---|---|
| 2033231-005 SPARE EASYONE-CS SCREW SET | <p>Screw set, consisting of:</p> <ul style="list-style-type: none"> - #1 Spring lock ring 12mm - #2 Torx screw K30x6, WN1452 - #2 Washer - #2 Allen screw M3x25 - #2 Hook-sleeve 5.6x6.5  |
| 2033231-006 SPARE EASYONE-CS BOARD WITH SENSOR | <ul style="list-style-type: none"> - PCB board with sensor and programmed serial number - Serial number label (Silver label)  |
| 2033231-007 SPARE EASYONE-CS LITHIUM BATTERY | BR1225 Lithium battery coin  |


7.4 Exchange Unit

| Part Number Short Description | Description |
|--|--|
| 2030431-002 RCDR EASYONE-CS | EasyOne-CS unit for exchange  |

8 Checkout Procedure

After repair or maintenance of the EasyOne-CS the following procedure is recommended in order to check correct operations of EasyOne-CS:

- Optional step: If all data in EasyOne-CS must be deleted, perform the ‘hidden key’ sequence 1-2-5 described in the following chapter in order to perform a factory reset. Please be aware that this step clears all customer data and resets the device to factory default settings.
- Get a calibration adapter and a calibration syringe (2030431-010)
- Ensure that the correct syringe volume is entered in the instrument’s configuration setting.
- Choose item “Check calibration” in the main menu of EasyOne.
- Connect the spirometer as shown below using the calibration adapter and the syringe. Ensure that the piston is fully inserted and at the stop position.



- Now press ENTER.
- Wait until the baseline has been set and you hear an audible signal.
- Now execute one full inspiratory pump stroke followed by one full expiratory pump stroke at moderate speed.
- After you perform the maneuver, you will see the text "Accuracy confirmed" at the top of the display and, beneath it, the percentage deviation and the average flow velocity of the pump stroke.
- If desired, make a printout of the calibration report.

9 Hidden Keys

The “hidden key” sequences access special features that are not available in the standard menu structure of EasyOne-CS. The following table shows what key sequences are available. The column labeled ‘Location’ indicates in what menu / screen the described key sequence has to be used.

Caution: The key sequences should not be used in conjunction with patient testing.

| Location | Key sequence | Description |
|---|--------------|---|
| Main Menu | 1-3-5 | Factory defaults (same as in Configuration Menu). |
| Device self test error screen | 1-3-5 | Ignore error and go on. Use only for troubleshooting! Do not perform a patient test, because the results could be erroneous |
| Device info screen (from Configuration Menu). | 1-2-4 | Show additional device information. |
| Device info screen | 1-2-5 | Hard factory reset (configuration settings are reset and database is deleted (In case of synchronisation with EasyWare, please do not forget to synchronize EasyOne with new database.) |
| Device info screen | 1-2-6 | FVC filter disable temporarily (until next power cycle). |
| Device info screen | 1-2-7 | Switch off BTPS correction temporarily (until next power cycle). |
| Device info screen | 1-2-8 | Switch off flow linearity correction temporarily (until next power cycle). |
| Device info screen | 1-3-9 | Restore corrupt database: Caution use only for inconsistent databases |

10 Device Error Codes

Internal errors are reported using the following screen:

ERROR
Device Selftest
Failure #xx
Please contact
service

The number xx indicates the type of error that is detected. The following list shows all error numbers and a description of the internal error.

| No. | Description |
|-----|---|
| 1 | internal software error (wrong argument) |
| 2 | internal software error (unsupported) |
| 3 | internal software error (timeout) |
| 4 | internal software error (data inconsistent) |
| 5 | internal software error (write error) |
| 6 | internal software error (read error) |
| 7 | internal software error (not found error) |
| 13 | internal software error (WD timeout) |
| 14 | the ultrasonic receive amplitude is too low (check spirette insertion) |
| 15 | the ultrasonic receive amplitude is too high (check spirette insertion) |
| 16 | Bootloader software inconsistent (CRC failure) |
| 17 | User software inconsistent (CRC failure). Update EasyOne-CS device software. |
| 18 | Calibration data inconsistent. |
| 19 | Failure of flash memory. |
| 20 | Flow sensor transit time range check error. Possible reasons: a) Blocked sound path, e.g. caused by wrongly inserted spirette, b) almost empty batteries. |
| 21 | ndd ASIC serial transmission error (STX expected) |
| 22 | ndd ASIC serial transmission error (sequence error) |
| 23 | ndd ASIC serial transmission error (unknown character) |
| 24 | Real time clock RAM error |
| 25 | Configuration data inconsistent (lithium battery failure possible) |
| 49 | internal software error (overrun error) |
| 50 | internal software error (timeout error) |
| 51 | internal software error (storage error) |
| 52 | internal software error (record structure error) |
| 53 | internal software error (address error) |
| 54 | Serial data flash or serial communication data inconsistent (CRC failure) |
| 55 | internal software error. Writing to the internal Flash failed (flash command sequence error) |
| 56 | internal software error. Erasing the internal Flash failed (flash block erase error) |
| 57 | internal software error. Writing to the internal Flash failed (flash program error) |
| 65 | internal software error. Writing to the internal Flash failed (flash block program error) |
| 66 | internal software error. Accessing the internal Flash failed (flash busy) |
| 67 | First record of database reached. Go beyond this record failed (FIRST_RECORD) |
| 68 | Last record of database reached. Go beyond this record failed (LAST_RECORD) |
| 69 | No valid record in the database found (NO_VALID_RECORD) |
| 70 | internal software error (UNKNOWN_ERROR) |
| 144 | Sensor serial communication error |

11 EasyOne-CS Device Software Update

11.1 Introduction

The internal software of EasyOne-CS is stored in a Flash EPROM. It is therefore possible to update the internal software. The following procedure describes how this update is executed.

11.2 Validated Upgrade Steps

The following upgrade steps are validated:

| Old Version | New Version | Item Number | EasyOne-CS device software | CRC |
|--------------|--------------|-------------|---|--------|
| V04.01.00.00 | V04.03.00.00 | 2033231-008 | GE-EOV04030000-P-EU_Resources.hex with EU language package | 0xF3D7 |

11.3 Preparation

The following items are required:

- 2033231-008.zip (upgrade package with EasyFlash and EasyOne-CS device software V4.3)
- ndd Screen-Connector cable 2030431-003.
- PC program EasyFlash, V2.4.12 or higher.
- EasyOne-CS device software for the download (*.hex).
- PC with Windows 2000, XP or Vista (all 32 Bit).

The following installation description presumes that you have basic knowledge of your PC operating system.

11.4 Unpacking of Upgrade Package

After getting upgrade package 2033231-008.zip the zip archive can be extracted to a folder of your choice, e.g. to C:\nddmed\2033231-008.

11.5 Installation of EasyFlash PC Software

The PC software EasyFlash is needed to update EasyOne-CS device software:

1. Start “C:\nddmed\2033231-008\EasyFlash\Setup.exe” to install the EasyFlash software.




2. Click “Next >”



3. Select target installation folder if you prefer another one. Click “Next >”.



4. Click “Next >”




5. The EasyFlash software is now installed on your PC. Click “Close” to terminate the setup program.
6. To start EasyFlash you find an appropriate short cut in the start menu of your PC.

11.6 EasyOne-CS Device Software Update Procedure

In order to update the EasyOne-CS device software follow this procedure:

1. Connect the EasyOne-CS Screen-Connector with an unused USB port of your PC.
2. Turn the power of EasyOne-CS on and connect it with the ScreenConnector. The EasyOne-CS display shows: „EasyOne“.
3. Start the PC software “EasyFlash”.
4. Set “Device Selection” to “EasyOne” and communication port to “Screen Connector”.
5. Press the “Update Device Info” button. The window on the right side shows the Model and the version information of the EasyOne-CS.
6. The device software can be selected through selecting the “Open” icon or with menu command “File/Open” C:\nddmed\2033231-008\ EasyOne-CS V4.3 Firmware EU\GE-EOV04030000-P-EU_Resources.hex”.



7. The programming is started by pressing the “Download” button. You are requested to enter your company and user name.
8. Under certain circumstances the following message can appear.




In order to proceed perform the following steps.

- a. Click “OK”.
- b. Disconnect EasyOne-CS from the ScreenConnector. The display of EasyOne-CS shows “Receive Idle mode, Setting 19200 Baud”.
- c. Reconnect EasyOne-CS with the ScreenConnector
- d. Click “Update Device Info” and press the “Download” button again.
9. During the programming of EasyOne-CS a progress bar is shown. The programming takes approx. 3 to 5 minutes. Please do not interrupt the programming process. After successful programming the following message is shown:



10. The EasyOne-CS device has to be restarted using the ON/OFF key.

11. The correct programming can be checked by pressing “Update Device Info” button and validating “Firmware Version” and “CRC” information.



Troubleshooting:

- If the programming of the software is interrupted: Turn the power of EasyOne-CS off and turn it on again. Retry the programming procedure.

12 EasyWare PC Software Installation

You will require Administrator privileges to install the software. Proceed as follows:


- Do not yet** connect the ScreenConnector to the PC.
- You will require Administrator privileges on the PC in order to install the software.
- Quit all programs and insert the EasyWare CD into your CD ROM drive. The installation routine should start automatically within 30 seconds of inserting the CD. If the installation routine does not start automatically, switch to root directory on the CD-ROM and run the "SmartInstallation.exe" program. You will then see the following window:




- Click on "Install EasyWare".




- Click on "Next >".



- In the bottom part of the window, you can still select whether this application is to be installed for all users ("Everyone") or for you only ("Just me").
- Click on "Next >" to confirm the displayed installation path. If you wish to change the displayed installation path, you can do this by clicking on "Browse..."



- Confirm the summarized installation information by clicking on "Next >".



- If button "Install FlashPlayer" is not visible the Macromedia FlashPlayer has also been installed automatically. Confirm completion of installation by clicking on "Close".
- Optional: On Windows Vista the FlashPlayer has to be installed manually (Macromedia Flash Player must be installed on your system in order to display the pediatric animation). Click on "Install FlashPlayer" if button is displayed.



- Confirm installation by clicking on "Yes".



- Click on "OK".

- Your computer may need to be restarted during this installation procedure under certain circumstances.
- The SmartInstaller can be terminated by clicking on "Exit". Remove the CD from the drive.
- An EasyWare icon will have been installed on your desktop after installation.
- Launch EasyWare. When EasyWare is launched for the first time, you are prompted to choose the language and enter the software serial number. Enter the software serial number that is provided on the CD-ROM sleeve and then quit EasyWare.

13 EasyOne-CS Lithium Battery Replacement

This procedure is intended for use by trained electronic service technicians. It describes how to exchange the internal Lithium battery.

13.1 Background

The internal Lithium battery of the EasOne-CS has a lifetime of more than 5 years (designed to last up to 10 years). If the lithium battery is empty, the device shows an error #25 message. The battery is needed to hold the date & time and parts of the device configuration information which is stored in the RAM area of the RTC. If the battery is empty, this data is corrupt and leads to the above error. In this case, it is necessary to change this battery.

13.2 Environment

1. Windows PC with installed and working EasyOne-CS connection (Serial cradle, USB cradle or Screen-Connector)
2. Installed EasyWare greater or equal V02.04
3. Torx T8 / Allen key /hex key screw driver depending on serial number
4. Electrostatic save working environment for the device disassembly


13.3 Preparation (save database)

To make sure that no data is lost/ altered, the internal EasyOne-CS database should be saved first. Please follow the below steps.


1. Start EasyWare by the link on the desktop
2. Switch EasyOne-CS on and press key sequence 1-3-5 when the below shown error screen appears. Enter the date & time if requested. Then the normal main menu should be shown.



3. Connect the EasyOne-CS to the PC (Device Connected is shown in the status line of EasyWare/ "PC interface active" on the EasyOne-CS screen)
4. Go to menu: File/Backup Device DB... And the following progress bar appears



5. The EasyOne-CS backup file will be written (takes several minutes) and the success is indicated by the following dialog




6. The backup file of the form EZ1BackupXXXXX_DateTime.txt
 - XXXXX: device serial number
 - Date: YearMonthDay
 - Time: HourMinute(e.g. **EZ1Backup40126_200506071607.txt**) is located in the folder where EasyWare has been started e.g. C:\programmfiles\ndd Medizintechnik\EasyWare


13.4 Battery exchange procedure

Please make sure that your working environment is electrostatic safe to avoid damage to the used electrostatic sensitive electronic components

- Open the front cover by removing the two main screws.
- Remove the printed circuit board (PCB)
- Do not remove the sensor tube (two additional screws)**
- Remove the Lithium battery (see figures below)



- Insert only 3V BR 1225 batteries which are approved by ndd Medizintechnik AG (e.g. Panasonic 3V BR 1225)
- Make sure that the + Sign form holder and battery are matched.



- Enter the new battery by preventing any short circuit (use only your fingers)
- Insert the PCB back to the EasyOne-CS case.
- Attach the front half of the case (make sure no cable will be squeezed)



13.5 Restart EasyOne-CS

Switch the EasyOne-CS back on. The above described error screen (#25) should be shown. Do the steps as described before. If you are in the main menu, check that all records are there by going to View results. If you are sure that all records are there, the exchange is finished.

13.6 Restore database

Under normal conditions EasyOne-CS restores the database by it self (Please wait... screen after 1-3-5). Double-check the amount of stored records. If the “Please wait...” screen appears longer than ca. 3 minutes you can interrupt it by pressing the ESC key (hold 0 key for more than 2sec.). This is only possible for firmware versions \geq V02.14.00.00. If restore database failed, it will be necessary to restore the database manually. **This can only be done at ndd!**

14 EasyOne-CS board with sensor and housing parts Replacement

This procedure is intended for use by trained electronic service technicians. It describes how to exchange the internal board with sensor and how to change the housing parts.

14.1 Background

If the sensors and/or the board fail, the complete board with the sensors has to be changed. If one housing part gets disturbed, change it following the description.

Please make sure that your working environment is electrostatic safe to avoid damage to the used electrostatic sensitive electronic components

- Open the battery door and remove the 2 AA batteries
- Open the front cover by removing the two main screws on the backside




- Remove the board (no screws) so you can remove the sensor tube by removing the two screws



- When you only change the front you must not remove the board with the sensors, remove the keypad from the damaged front and put it into the new front part.



- When you change the board, remove the old one and replace it with the new one. Bring the sensor tube into position and screw it with the washers – be carefully, do not turn the screws too tight! Bring the board into position.



- Put the front part with the keyboard back on the instrument in that way that the window is placed over the LCD display and that the front fits tight on the back cover
- Turn the instruments (front at the bottom) and turn the screws in. The holes in the hook sleeves must be in the right position, look to the left and right side of the instrument



- When you change the rear half of the housing, please remove the pocket and the serial number label (silver label) from the old housing part and put the pocket into the rear half and glue the label at the same position on the new rear half part.





Pocket


- When you change the board with sensor, remove the old label and glue the new delivered one onto the rear half.
- Put in the 2 AA batteries, close the battery door


15 Appendix


15.1 Schematic Diagrams

















15.2 WEEE Selective Treatment

WEEE category: 8 (Medical Device)

| WEEE Selective Treatment Category <i>Traitements sélectifs des matières et composants des DEEE</i> | Type/Materials <i>Substances, préparations et composants</i> | Check ONLY if substance is present <i>Une croix si la substance est présente</i> |
|--|--|--|
| 1 | Capacitors containing Polychlorinated biphenyls (PCB) or Polychlorinated terphenyls (PCY/PCT) <i>Condensateurs contenant du polychlorobiphényle (PCB), et des polychlorotérphényles (PCY)</i> | |
| 2 | Components containing Mercury - e.g. switches, circuit breakers, lamps <i>Composants contenant du mercure, tels que les interrupteurs ou les lampes à rétroéclairage</i> | |
| 3 | Batteries - <i>Piles et accumulateurs</i> | X 2033231-007 |
| 4 | Printed Circuit Boards - Area greater than 10 square centimeters <i>Cartes de circuits imprimés si la surface est supérieure à 10 centimètres carrés</i> | X 2033231-006 |
| 5 | Toner cartridges from printers - <i>Cartouches de toner</i> | |
| 6 | Plastic containing brominated flame retardants <i>Matières plastiques contenant des retardateurs de flamme bromés</i> | |
| 7 | Components containing Asbestos <i>Déchets d'amiante et composants contenant de l'amiante</i> | |
| 8 | Cathode ray tubes - <i>Tubes cathodiques</i> | |
| 9 | Refrigerants - Chlorofluorocarbons (CFC), Hydrochlorofluorocarbons (HCFC) or hydrofluorocarbons (HFC), and hydrocarbons (HC), <i>Réfrigérants - Chlorofluorocarbones (CFC), hydrochlorofluorocarbone (HCFC) ou hydrofluorocarbone (HFC), hydrocarbures (HC)</i> | |
| 10 | discharge lamps - <i>Lampes à décharge</i> | |
| 11 | Liquid crystal displays (together with their casing where appropriate) of a surface area greater than 100 square centimeters and all back-lighted with gas discharge lamps <i>Écrans à cristaux liquides (ainsi que leur boîtier le cas échéant) d'une surface sup</i> | |
| 12 | External electrical cables - <i>Câbles électriques extérieurs</i> | X 2030431-003 |
| 13 | Components containing refractory ceramic fibers <i>Composants contenant des fibres céramiques réfractaires</i> | |
| 14 | Components containing radioactive substances <i>Composants contenant des substances radioactives</i> | |
| 15 | Electrolyte capacitors containing substances of concern — greater than 25 mm diameter or height or similar volume <i>Condensateurs électrolytiques contenant des substances dangereuses (hauteur >25mm, diamètre >25mm ou volume proportionnellement similaire)</i> | |
| 16 | Lead - <i>Plomb</i> | |
| 17 | Cadmium - <i>Cadmium</i> | |
| 18 | Fluids - <i>Fluides</i> | |

GE Medical Systems
Information Technologies, Inc.
8200 West Tower Avenue
Milwaukee, WI 53233 USA
Tel: +1 414 355 5000
+1 800 437 1171 (USA only)
+1 800 668 0732 (Canada only)
Fax: +1 414 355 3790

GE Medical Systems
Information Technologies, GmbH
Munzinger Straße 5
D-79111 Freiburg
Germany
Tel: +49 761 45 43 - 0
Fax: +49 761 45 43 - 233

Asia Headquarters
GE China Co., Ltd.
No1 Huatuo Road,
Zhangjiang Hi-Tech Park Pudong,
Shanghai, P.R.China 201203
Tel: +86 21 38777888
Fax: +86 21 38777402